

CALIBRATION CERTIFICATE

for
Kustom LaserCam 4 Speed Meter

Customer:

Devon and Cornwall Police

Serial Number:

LE3318

Certificate Number:

8921

Date of Calibration:

09 April 2025

Next Calibration:

08 April 2026

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999915 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
- Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m

Actual

47.64 m

108.69 m

148.86 m

Indicated

47.6 m

108.7 m

148.9 m

Speed ± 1 MPH

Actual

72 MPH

-45 MPH

26 MPH

53 MPH

-64 MPH

Indicated

72 MPH

-45 MPH

26 MPH

53 MPH

-64 MPH


The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedometer detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korcz
Technical Signatory



TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: info@truvelouk.com